

# OLYMPUS®

Your Vision, Our Future

# VISERA ELITE II

VISERA ELITE II VIDEO SYSTEM CENTER

# OTV-S200

All-in-One 2D Processor and Light Source



# OTV-S200

## All-in-One 2D Processor and Light Source

- Capable of 2D observation
- Compact system for simplified work flow

## LCD Touch Panel

- Intuitive handling and setup
- Presets allow easy preparation and maintenance

## LED Light Source

- Reduction in running costs thanks to the long lifetime of the LEDs
- Excellent natural color reproduction with the combination of an enhanced imaging process

## Special Light Observation

- NBI and two modes of IR observation

Specifications			
<b>Power Supply</b>	Rated voltage	100–240 V AC; within $\pm 10\%$	
	Rated frequency	50/60 Hz; within $\pm 1$ Hz	
	Rated input	400 VA	
<b>Size</b>	Dimensions (maximum)	383 (W) $\times$ 199 (H) $\times$ 506 (D) mm	
	Weight	19.3 kg	
<b>Observation</b>	Analog signal output	VBS composite and Y/C; simultaneous outputs possible.	
	Digital signal output	HD-SDI (SMPTE292M), DVI (WUXGA, 1080 pixels, or SXGA can be selected)	
	Electronic zoom	The image enlargement level can be selected. 3 modes (1.0 $\times$ , 1.2 $\times$ , 1.5 $\times$ )	
	Optical-digital observation	The optical-digital observation can be performed. The endoscope compatible with the optical-digital observation is required.	
		NBI observation	This observation mode uses narrow-band light
IR observation	This observation mode uses infrared light		
<b>Documentation</b>	Remote control	The following ancillary equipment can be controlled (specified models only): · Portable memory / · Video recorder / · Video printer / · Image filing system	
	Recording format and number of recorded images in internal memory	TIFF: no compression	Approx. 120 images
		JPEG (1/5): Approx. 1/5 compression	Approx. 636 images
		JPEG (1/10): Approx. 1/10 compression	Approx. 1108 images
	These are the numbers of recorded images when both HDTV and SDTV images are recorded. These numbers vary depending on the images.		
<b>Illumination</b>	Examination lamp	LED	
	Cooling	Forced-air cooling	
	Providing maximum light intensity	Less than 5.64 W (approximately same as Olympus light source that is mounted xenon short-arc lamp 300 W)	
	Observation mode	WLI or NBI observation IR observation (when connected to CLV-S200-IR)	
<b>Automatic Brightness Adjustment</b>	Automatic brightness adjustment method	LED drive current control	
	Automatic exposure	17 steps	
	Brightness mode	Auto Manual	
<b>Classification (Medical Electrical Equipment)</b>	Type of protection against electric shock	Class I	
	Degree of protection against electric shock of applied part	Depends on applied part. Also refer to applied part (camera head or videoscope).	
	Degree of protection against explosion	The video system center should be kept away from flammable gases	

Specifications, design and accessories are subject to change without any notice or obligation on the part of the manufacturer.



OLYMPUS MEDICAL SYSTEMS CORP.  
Shinjuku Monolith, 2-3-1 Nishi-Shinjuku, Shinjuku-ku, Tokyo 163-0914, Japan

For a complete listing of sales and distribution locations visit:  
[www.olympus.com](http://www.olympus.com)